

Document No.: MDCK-BRP-ZL-07

Version No.: A/0

Residual MDCK DNA Quantitation KitProduct No.: **SK030209M100**Lot No.: **20260401**Date of Mfg.: **21 Apr. 2026**Expiration Date: **22 Oct. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	3 fg/ μ L MDCK DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or absence of specific amplification.	PASS
MDCK DNA Control	Report the results.	30.8 ng/ μ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiqingDAY/MONTH/YEAR 24/04/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China