

Document No.: Human-BRP-ZL-03

Version No.: A/3

**Residual Human DNA Quantitation Kit**Product No.: **SK030207H100**Lot No.: **20250401**Date of Mfg.: **11 Apr. 2025**Expiration Date: **04 Nov. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$ .	PASS
Repeatability	30 fg/ $\mu$ L Human DNA Control was detected, $CV \leq 15.0\%$ .	PASS
No Template Control (NTC)	The detection value of NTC should not be more than 6.00 fg/ $\mu$ L.	PASS

Approved By Steven chen DAY/MONTH/YEAR 17/04/2025  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China