

Document No.: HCP-BRP-ZL-103

Version No.: A/4

***E. coli* (Protein Expression Strains) HCP ELISA Kit
(One-step ELISA)**Product No.: **1301301-1**Lot No.: **20260301**Date of Mfg.: **03 Mar. 2026**Expiration Date: **13 Feb. 2027**

Test Item	Specification	Result
<i>E. coli</i> HCP-E Calibration Standard	Report the results.	0.70 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Quantitation Limit (QL)	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450 \text{ nm}}$) should not be more than 0.200.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 04/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China