

Document No.: *E. coli*-BRP-ZL-11

Version No.: A/4

Residual *E. coli* DNA Quantitation Kit (2G)

Product No.: 1101107-1

Lot No.: 20260301

Date of Mfg.: 02 Mar. 2026

Expiration Date: 03 Jan. 2028

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L <i>E. coli</i> DNA Control was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS
<i>E. coli</i> DNA Control	Report the results.	30.4 ng/ μ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiqingDAY/MONTH/YEAR 05/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China