

Document No.: Vero-BRP-ZL-10

Version No.: A/4

**Residual Vero DNA Size Analysis Kit (2G)**Product No.: **1103174**Lot No.: **20260302**Date of Mfg.: **03 Mar. 2026**Expiration Date: **01 Jul. 2027**

Test Item	Specification	Result	
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS	
Amplification efficiency of Standard curve	$E = 83.3\% - 110.0\%$ .	PASS	
Repeatability	Vero Primer&Probe MIX-85	30 fg/ $\mu$ L Vero DNA Control was detected, $CV \leq 20.0\%$ .	PASS
	Vero Primer&Probe MIX-134		PASS
	Vero Primer&Probe MIX-229		PASS
	Vero Primer&Probe MIX-552		PASS
No Template Control (NTC)	The Ct value of NTC should be 2 larger than the mean Ct value of the lowest concentration in the standard curve.	PASS	
Vero DNA Control	Report the results.	31.9 ng/ $\mu$ L	

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyangDAY/MONTH/YEAR 16/03/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China