

Document No.: HEK-BRP-ZL-07

Version No.: A/4

Residual HEK293 DNA Quantitation Kit (3G)

Product No.: 1101104-1

Lot No.: 20260301

Date of Mfg.: 03 Mar. 2026

Expiration Date: 09 Sep. 2027

| Test Item | Specification | Result |
|--|--|------------------|
| Linearity of Standard curve | $R^2 \geq 0.990$. | PASS |
| Amplification efficiency of Standard curve | $E = 90.0\% - 110.0\%$. | PASS |
| Repeatability | 30 fg/ μ L HEK293 DNA Control was detected, $CV \leq 15.0\%$. | PASS |
| No Template Control (NTC) | The mean value of NTC should not be more than 6.00 fg/ μ L. | PASS |
| HEK293 DNA Control | Report the results. | 30.9 ng/ μ L |

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyangDAY/MONTH/YEAR 10/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China