

Document No.: BSA-BRP-ZL-88

Version No.: A/2

**Bovine Serum Albumin (BSA)
ELISA Kit**Product No.: **1401401**Lot No.: **20260301**Date of Mfg.: **05 Mar. 2026**Expiration Date: **17 Dec. 2026**

Test Item	Specification	Result
BSA Calibration Standard	Report the results.	0.52 µg/mL
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	16 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Quantitation Limit (QL)	1 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.150.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 09/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China