

Document No.: RCL-BRP-ZL-03

Version No.: A/2

Replication-Competent Lentivirus (RCL) Quantitation KitProduct No.: **1403441**Lot No.: **20260301**Date of Mfg.: **05 Mar. 2026**Expiration Date: **05 Feb. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 83.3\% - 110.0\%$.	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles.	PASS
RCL Control	Report the results.	6.38×10^9 copies / μ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyangDAY/MONTH/YEAR 11/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China