

Document No.: Vero-BRP-ZL-03

Version No.: A/4

Residual Vero DNA Quantitation KitProduct No.: **SK030204V100**Lot No.: **20260302**Date of Mfg.: **04 Mar. 2026**Expiration Date: **27 Sep. 2027**

| Test Item | Specification | Result |
|--|---|------------------|
| Linearity of Standard curve | $R^2 \geq 0.990$. | PASS |
| Amplification efficiency of Standard curve | $E = 90.0\% - 110.0\%$. | PASS |
| Repeatability | 3 fg/ μ L Vero DNA Control was detected, $CV \leq 15.0\%$. | PASS |
| No Template Control (NTC) | The Ct value of NTC should not be less than 35.00 or absence of specific amplification. | PASS |
| Vero DNA Control | Report the results. | 31.4 ng/ μ L |

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyangDAY/MONTH/YEAR 17/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China