

Document No.: MVM-BRP-ZL-05

Version No.: A/1

MVM Detection KitProduct No.: **1506742**Lot No.: **20260301**Date of Mfg.: **10 Mar. 2026**Expiration Date: **22 Dec. 2027**

Test Item	Specification	Result
Low Concentration Sample	FAM channel: Ct < 40.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
No Template Control (NTC)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyiDAY/MONTH/YEAR 12/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China