

Document No.: BVDV-BRP-ZL-04

Version No.: A/0

BVDV Detection KitProduct No.: **1506731**Lot No.: **20260301**Date of Mfg.: **10 Mar. 2026**Expiration Date: **23 Dec. 2026**

Test Item	Specification	Result
Positive Control	High concentration sample was detected, FAM channel: Ct value < 25.00 with specific amplification curve, VIC channel: Ct value < 35.00 with specific amplification curve.	PASS
	Medium concentration sample was detected, FAM channel: Ct value < 32.00 with specific amplification curve, VIC channel: Ct value < 35.00 with specific amplification curve.	PASS
	Low concentration sample was detected, FAM channel: Ct value < 38.00 with specific amplification curve, VIC channel: Ct value < 35.00 with specific amplification curve.	PASS
No Template Control (NTC)	FAM channel: Ct value \geq 40.00 or absence of specific amplification curve. VIC channel: Ct value < 35.00 with specific amplification curve.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Steven ChenDAY/MONTH/YEAR 12/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China