

Document No.: CHO-K1 HCP-BRP-ZL-12

Version No.: A/1

**CHO-K1 HCP ELISA Kit
(One-step ELISA)**Product No.: **1301305-1**Lot No.: **20260301**Date of Mfg.: **10 Mar. 2026**Expiration Date: **13 Feb. 2027**

Test Item	Specification	Result
CHO-K1 HCP Calibration Standard	Report the results.	0.96 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	150 ng/mL sample solution was tested, $CV \leq 20.0\%$.	PASS
Quantitation Limit (QL)	3 ng/mL sample solution was tested, $CV \leq 25.0\%$, and the relative bias of the concentration should not be more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.200.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 13/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China