

Document No.: Pichia-BRP-ZL-01

Version No.: A/3

Residual *Pichia Pastoris* DNA Quantitation KitProduct No.: **SK030205P100**Lot No.: **20260301**Date of Mfg.: **10 Mar. 2026**Expiration Date: **11 Dec. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L <i>Pichia Pastoris</i> DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or absence of specific amplification.	PASS
<i>Pichia Pastoris</i> DNA Control	Report the results.	30.1 ng/ μ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiqingDAY/MONTH/YEAR 17/03/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China