

Document No.: Virus-BRP-ZL-49

Version No.: A/1

**MVM Positive Control**Product No.: **1506747**Lot No.: **20260301**Date of Mfg.: **17 Mar. 2026**Expiration Date: **11 Aug. 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeijiangDAY/MONTH/YEAR 18/03/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China