

Document No.: SMNE-BRP-ZL-29

Version No.: A/1

**Nonspecific Endonuclease (SMNE) ELISA Kit**Product No.: **1402421**Lot No.: **20260301**Date of Mfg.: **24 Mar. 2026**Expiration Date: **13 Jul. 2026**

Test Item	Specification	Result
SMNE Calibration Standard	Report the results.	5.1 µg/mL
Linearity of Standard Curve	$R^2 \geq 0.990$ .	PASS
Repeatability	10 ng/mL sample solution was tested, $CV \leq 15.0\%$ .	PASS
Quantitation Limit (QL)	0.25 ng/mL sample solution was tested, the CV and relative bias of the concentration were not more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{nm}}$ ) should not be more than 0.150.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JinqunDAY/MONTH/YEAR 24/03/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China