

Document No.: Virus-BRP-ZL-46

Version No.: A/1

Vesivirus 2117 Detection KitProduct No.: **1506743**Lot No.: **20260301**Date of Mfg.: **23 Mar. 2026**Expiration Date: **26 Feb. 2027**

Test Item	Specification	Result
Low concentration sample	FAM channel: Ct < 40.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	Pass
No Template Control (NTC)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	Pass

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyiDAY/MONTH/YEAR 01/04/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China