

MVM Positive Control

Product No.: 1506747

(For Research Use Only)

■ Product description

SHENTEK® MVM Positive Control is intended for use with SHENTEK® Virus DNA & RNA Extraction and Detection Kits, and serves as a positive control to monitor the entire process of MVM DNA extraction and PCR detection assay.

■ Kit contents and storage

Reagent	Quantity	Storage
MVM Positive Control	100 µL × 3 tubes	below -65°C

MVM Positive Control can be stored at appropriate conditions for up to 12 months. Please check the expiration date on the labels.

■ Attentions

1. Thaw one tube completely at 2-8°C or on ice, then vortex and spin briefly.
2. Follow the positive control sample (PCS) preparation protocol outlined in the user guide of SHENTEK® MVM Detection Kit.
3. This product is ready for direct use without dilution.
4. Use only after it is equilibrated to room temperature, and avoid repeated freeze-thaw cycles.
5. Do not mix reagents from different batches.
6. Do not use reagents beyond their expiration date.
7. This product is intended for research use only and is not used for clinical diagnosis or treatment. Use with appropriate caution.
8. This product is an inactivated replication-defective pseudovirus, which is safe, but it is still recommended to operate it in a biosafety cabinet. Experimental waste generated during its use must be sterilized via autoclaving and disposed of in compliance with medical waste regulations.

■ **Related products**

Product Name	Product No.	Quantity
SHENTEK® Virus DNA & RNA Extraction Kit	1506730	50 Extractions
SHENTEK® MVM Detection Kit	1506742	50 Reactions

Effective date: 11 Dec. 2024

Support & Contact

SHENTEK

Huzhou Shenke Biotechnology Co., Ltd.
www.shentekbio.com
Address: 8th Floor, 6B Building, No.1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China
E-mail: info@shentekbio.com
Phone: +1 (908) 822-3199 / (+86) 400-878-2189