

Document No.: HCP-BRP-ZL-19

Version No.: A/3

***E. coli* (Protein Expression Strains) HCP ELISA Kit**

Product No.: 1301301

Lot No.: 20260601

Date of Mfg.: 25 Jun. 2026

Expiration Date: 24 Jun. 2027

Test Item	Specification	Result
<i>E.coli</i> HCP-E Calibration Standard	Report the results.	0.78 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Quantitation Limit	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias within $\pm 20.0\%$.	PASS
Negative Control Sample (NCS)	The mean value of NCS (OD _{450 nm}) should not be more than 0.150.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 01/07/2026

Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China