

Document No.: Reverse Transcriptase-BRP-ZL-02

Version No.: A/2

Reverse Transcriptase Positive ControlProduct No.: **1505701**Lot No.: **20260601**Date of Mfg.: **23 Jun. 2026**Expiration Date: **18 May 2027**

Test Item	Specification	Result
Positive Control Sample (PCS)	The Ct value of PCS should be not more than 28.00.	PASS
Negative Control Sample (NCS)	The Ct value of NCS (Buffer A) should be not less than 40.00.	PASS
For Research Use Only. Not for diagnostic or therapeutic use.		

Approved By Zong WeiyiDAY/MONTH/YEAR 29/06/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China