

Document No.: *P.pastoris* HCP-BRP-ZL-06

Version No.: A/3

***P.pastoris* HCP ELISA Kit**
(One-step ELISA)

Product No.: 1301313

Lot No.: 20260601

Date of Mfg.: 30 Jun. 2026

Expiration Date: 29 Jun. 2027

Test Item	Specification	Result
<i>P.pastoris</i> HCP Calibration Standard	Report the results.	1.2 µg/bottle
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Repeatability	50 ng/mL sample solution was tested, $CV \leq 20.0\%$.	PASS
Quantitation Limit (QL)	2 ng/mL sample solution was tested, $CV \leq 25.0\%$, relative bias within $\pm 25.0\%$.	PASS
Negative Control Sample (NCS)	The mean value of NCS (OD _{450 nm}) should not be more than 0.300.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 02/07/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China