

Document No.: Virus-BRP-ZL-48

Version No.: A/1

Vesivirus 2117 Positive ControlProduct No.: **1506743-R01**Lot No.: **20260701**Date of Mfg.: **24 Jul. 2026**Expiration Date: **16 Jul. 2027**

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyiDAY/MONTH/YEAR 29/07/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China