

Document No.: rcAAV-BRP-ZL-11

Version No.: A/0

rcAAV-5/N Quantitation Kit

Product No.: 1403445-1

Lot No.: 20260501

Date of Mfg.: 11 May 2026

Expiration Date: 29 Jul. 2027

Test Item	Specification		Result
Linearity of Standard curve	Target	$R^2 \geq 0.990$.	PASS
	Reference		
Amplification efficiency of Standard curve	Target	$E = 83.3\% - 110.0\%$.	PASS
	Reference		
Repeatability	Target	Low concentration sample was detected, $CV \leq 20.0\%$.	PASS
	Reference		
No Template Control (NTC)	The Ct value of NTC should be higher than the mean Ct value of the quantitation limit (QL) concentration.		PASS
T&R DNA Control-5	Report the results.		4.68×10^9 copies/ μ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong Weiping

DAY/MONTH/YEAR 21/05/2026
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China