

Document No.: CHO-K1 HCP-BRP-ZL-12

Version No.: A/1

## CHO-K1 HCP ELISA Kit (One-step ELISA)

Product No.: **1301305-1**Lot No.: **20260502**Date of Mfg.: **25 May 2026**Expiration Date: **13 Feb. 2027**

Test Item	Specification	Result
CHO-K1 HCP Calibration Standard	Report the results.	0.96 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$ .	PASS
Repeatability	150 ng/mL sample solution was tested, the CV should not be more than 20.0%.	PASS
Quantitation Limit (QL)	3 ng/mL sample solution was tested, the CV and relative bias of the concentration should not be more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS (OD <sub>450 nm</sub> ) should not be more than 0.200.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 02/06/2026  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)

For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China