

Document No.: MP-BRP-ZL-19

Version No.: A/1

AdvSHENTEK Mycoplasma DetectInnova Cassette

Product No.: **16031010405**

Lot No.: **20260501**

Size: **4 Cassettes /Kit**

Date of Mfg.: **28 May 2026**

Expiration Date: **18 Mar. 2027**

Appearance and integrity	Specification	Result
Visual Inspection	The exterior of the Mycoplasma DetectInnova cassette is clean, smooth, and free of burrs and stains.	PASS
	The sealing film of cassette must be intact, with no signs of leakage.	PASS
	All reagent components in bottles and tubes are free of leakage.	PASS
Label/QR code Inspection	The label is complete, legible, and accurate.	PASS
	QR Code is scannable and links to the corresponding program.	PASS
Test Item	Specification	Result
Positive Control Detection	<i>Mycoplasma hyorhinitis</i> (10 CFU/mL) detection, Result: Positive.	PASS
	<i>Mycoplasma orale</i> (10 CFU/mL) detection, Result: Positive.	PASS
	<i>Mycoplasma pneumoniae</i> (10 CFU/mL) detection, Result: Positive.	PASS
Negative Control Detection	Result: Negative (No detection observed)	PASS
Repeatability	Detection of MP Positive Control (Part No.:NNA068) shows CV ≤ 10% for both FAM and ROX signals.	PASS
Packaging and Stability	Specification	Result
Packaging	The product packaging is intact with no damage or leakage, and the quantity is correct.	PASS
Stability	The product is stable for 12 months when stored at 2–8 °C.	PASS

Certificate of Analysis

Statement	
This batch has undergone comprehensive testing and meets the company's quality standards.	
Conclusion	
Batch testing conclusion	PASS
Note	This COA is only applicable to the above-mentioned lot number products. The detection method is based on an internally validated detection program.
For Research Use Only. Not for diagnostic or therapeutic use.	

Approved By Zong WeiyiDAY/MONTH/YEAR 29/05/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China