

Document No.: BHK-BRP-ZL-07

Version No.: A/1

**Residual BHK DNA Quantitation Kit**Product No.: **1101116**Lot No.: **20260602**Date of Mfg.: **03 Jun. 2026**Expiration Date: **09 Jul. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$ .	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$ .	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or absence of specific amplification.	PASS
BHK DNA Control	Report the results.	30.5 ng/ $\mu$ L

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiqingDAY/MONTH/YEAR 11/06/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China