

Document No.: 293T HCP-BRP-ZL-22

Version No.: A/2

**HEK293 HCP ELISA Kit  
(One-step ELISA)**Product No.: **1301311Z01**Lot No.: **20260601**Date of Mfg.: **03 Jun. 2026**Expiration Date: **02 Jun. 2027**

Test Item	Specification	Result
HEK293 HCP Calibration Standard	Report the results.	0.59 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$ .	PASS
Repeatability	135 ng/mL sample solution was tested, $CV \leq 20.0\%$ .	PASS
Quantitation Limit (QL)	6 ng/mL sample solution was tested, $CV \leq 25.0\%$ , relative bias of concentration should not be more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{ nm}}$ ) should not be more than 0.200.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Cheng JiajunDAY/MONTH/YEAR 12/06/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China