

Document No.: Fun-BRP-ZL-13

Version No.: A/1

**Fungi DNA Detection Kit**Product No.: **1504631**Lot No.: **20260201**Date of Mfg.: **13 Feb. 2026**Expiration Date: **01 Dec. 2027**

Test Item	Specification	Result
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Ct value of FAM channel $\geq$ 39.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

For Research Use Only. Not for diagnostic or therapeutic use.

Approved By Zong WeiyangDAY/MONTH/YEAR 27/02/2026  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China