

Certificate of Analysis

Document No.: HCP-BRP-ZL-88

Version No.: A/0

Bovine Serum Albumin (BSA) ELISA Kit

Product No.: 1401401

Lot No.: 20250301

Date of Mfg.: 13 Mar. 2025

Expiration Date: 05 Nov. 2025

Test Item	Specification	Result
Linearity of Standard curve	R²≥0.990.	PASS
Repeatability	16 ng/mL sample solution was tested, CV≤15.0%.	PASS
Limit of Quantification	1 ng/mL sample solution was tested, CV≤20.0%, relative bias of concentration was not more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450 \text{ nm}}$) should be not more than 0.150.	PASS

Approved By Long Weiging

DAY/MONTH/YEAR 4/03/XX
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China