

Document No.: CAR-BRP-ZL-03

Version No.: A/1

## CAR/TCR Transgene Copy Number Quantitation Kit

Product No.: SK030221CA100

Lot No.: 20250401

Date of Mfg.: 27 Apr. 2025

Expiration Date: 28 Oct. 2026

Test Item	Specification		Result
Linearity of Standard curve	CAR/TCR	R <sup>2</sup> ≥0.990.	PASS
	SCG		
Amplification efficiency of Standard curve	CAR/TCR	E=90.0%-110.0%.	PASS
	SCG		
Repeatability	CAR/TCR	2.33×10 <sup>2</sup> copies/μL CAR/TCR Control was detected, CV≤15.0%.	PASS
	SCG		
No Template Control (NTC)	The Ct value of NTC should be larger than the mean Ct value of the lowest concentration in the standard curve.		PASS

Approved By Zong WeiqingDAY/MONTH/YEAR 06/05/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China