

Document No.: CHO HCP-BRP-ZL-28

Version No.: A/3

**CHO HCP ELISA Kit  
(One-step ELISA)**Product No.: **1301304-1**Lot No.: **20250401**Date of Mfg.: **27 Apr. 2025**Expiration Date: **26 Feb. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Repeatability	16 ng/mL sample solution was tested, $CV \leq 15.0\%$ .	PASS
Limit of Quantification	1 ng/mL sample solution was tested, $CV \leq 25.0\%$ , relative bias of concentration was not more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{ nm}}$ ) should not be more than 0.200.	PASS

Approved By Zong WeijiangDAY/MONTH/YEAR 09/05/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China