

Document No.: CHO HCP-BRP-ZL-28

Version No.: A/3

CHO HCP ELISA Kit (One-step ELISA)

Product No.: **1301304-1**Lot No.: **20250802**Date of Mfg.: **20 Aug. 2025**Expiration Date: **19 Aug. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Repeatability	16 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Limit of Quantification	1 ng/mL sample solution was tested, $CV \leq 25.0\%$, relative bias of concentration was not more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{nm}}$) should not be more than 0.200.	PASS

Approved By Cheng JiajumDAY/MONTH/YEAR 26/08/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China