

Document No.: CHO-K1 HCP-BRP-ZL-03

Version No.: A/0

CHO K1 HCP ELISA Kit
(One-step ELISA)

Product No.: 1301305

Lot No.: 20250101

Date of Mfg.: 03 Jan. 2025

Expiration Date: 06 Nov. 2025

Quality Control Data:

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	40 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Limit of Quantitation	3 ng/mL sample solution was tested, the CV and relative bias of the concentration were no more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}-620\text{ nm}}$) should be no more than 0.300.	PASS

Quality Assurance

Zong Weiping

DAY/MONTH/YEAR

2025/01/09

Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China