

Document No.: CHO-K1 HCP-BRP-ZL-12

Version No.: A/0

**CHO-K1 HCP ELISA Kit
(One-step ELISA)**Product No.: **1301305-1**Lot No.: **20250401**Date of Mfg.: **14 Apr. 2025**Expiration Date: **23 Mar. 2026**

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	150 ng/mL sample solution was tested, the CV should not be more than 20.0%.	PASS
Limit of Quantification	3 ng/mL sample solution was tested, the CV and relative bias of the concentration should not be more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS (OD _{450 nm}) should not be more than 0.200.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 21/04/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China