

Document No.: HCP-BRP-ZL-88

Version No.: A/1

**Bovine Serum Albumin (BSA)  
ELISA Kit**Product No.: **1401401**Lot No.: **20251201**Date of Mfg.: **04 Dec. 2025**Expiration Date: **02 Apr. 2026**

Test Item	Specification	Result
BSA Calibration standard	Report the results.	0.50 µg/mL
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Repeatability	16 ng/mL sample solution was tested, $CV \leq 15.0\%$ .	PASS
Quantitation Limit	1 ng/mL sample solution was tested, $CV \leq 20.0\%$ , relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{ nm}}$ ) should not be more than 0.150.	PASS

Approved By Cheng JinqinDAY/MONTH/YEAR 08/12/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China