

Document No.: HCP-BRP-ZL-19

Version No.: A/2

***E. coli* (Protein Expression Strains) HCP ELISA Kit**

Product No.: 1301301

Lot No.: 20260102

Date of Mfg.: 13 Jan. 2026

Expiration Date: 12 Jan. 2027

Test Item	Specification	Result
<i>E. coli</i> HCP-E Calibration Standard	Report the results.	0.75 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Quantitation Limit	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.150.	PASS

Approved By Cheng JiaqianDAY/MONTH/YEAR 13/01/2026

Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China