

Document No.: Fun-BRP-ZL-13

Version No.: A/0

Fungi DNA Detection KitProduct No.: **1504631**Lot No.: **20250902**Date of Mfg.: **22 Sep. 2025**Expiration Date: **22 May 2027**

Test Item	Specification	Result
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Ct value of FAM channel ≥ 39.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 25/09/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China