

Document No.: 293T HCP-BRP-ZL-22

Version No.: A/0

HEK293 HCP ELISA Kit (One-step ELISA)

Product No.: 1301311Z01

Lot No.: 20250901

Date of Mfg.: 02 Sep. 2025

Expiration Date: 01 Sep. 2026

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	135 ng/mL sample solution was tested, $CV \leq 20.0\%$.	PASS
Limit of Quantification	6 ng/mL sample solution was tested, $CV \leq 25.0\%$, relative bias of concentration should not be more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.200.	PASS

Approved By Chang Jiajia

DAY/MONTH/YEAR 02/09/2025
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China