

Document No.: MVM-BRP-ZL-05

Version No.: A/0

MVM Detection KitProduct No.: **1506742**Lot No.: **20251201**Date of Mfg.: **10 Dec. 2025**Expiration Date: **24 Dec. 2026**

Test Item	Specification	Result
Low concentration sample	FAM channel: Ct < 40.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
No Template Control (NTC)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

Approved By Zheng LipingDAY/MONTH/YEAR 10/12/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China