

Document No.: Mycop-BRP-ZL-26

Version No.: A/2

**Mycoplasma DNA Detection Kit (2G)**

Product No.: 1509841

Lot No.: 20251101

Date of Mfg.: 04 Nov. 2025

Expiration Date: 03 Jun. 2027

Test Item	Specification	Result
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	Pass
No Template Control (NTC)	Ct value of FAM channel $\geq$ 40.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	Pass

Approved By Zheng LipingDAY/MONTH/YEAR 05/11/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China