

Document No.: Virus-BRP-ZL-74

Version No.: A/0

PCV-1 Detection KitProduct No.: **1506744**Lot No.: **20251101**Date of Mfg.: **19 Nov. 2025**Expiration Date: **13 Oct. 2027**

Test Item	Specification	Result
Sensitivity	Low concentration sample was detected, Ct value of FAM channel < 40.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Ct value of FAM channel undetected or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 21/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China