

Document No.: Virus-BRP-ZL-75

Version No.: A/0

PCV-1 Positive ControlProduct No.: **1506744-R01**Lot No.: **20251101**Date of Mfg.: **19 Nov. 2025**Expiration Date: **14 Oct. 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	The Ct value of PCS should be not more than 35.00.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be undetected or absence of specific amplification.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 21/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China