

Document No.: Virus-BRP-ZL-75

Version No.: A/0

### PCV-1 Positive Control

Product No.: **1506744-R01**

Lot No.: **20260101**

Date of Mfg.: **19 Jan. 2026**

Expiration Date: **14 Oct. 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	The Ct value of PCS should be not more than 35.00.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be undetected or absence of specific amplification.	PASS

Approved By Zong Weiqing

DAY/MONTH/YEAR 26/01/2026  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)

For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China