

Document No.: HCP-BRP-ZL-63

Version No.: A/0

Porcine Trypsin ELISA KitProduct No.: **1402422**Lot No.: **20251101**Date of Mfg.: **25 Nov. 2025**Expiration Date: **18 Aug. 2026**

Test Item	Specification	Result
Trypsin Calibration Standard	Report the results.	46.6 ng/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	1.28 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Quantitation Limit	0.04 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.200.	PASS

Approved By Cheng JiajunDAY/MONTH/YEAR 26/11/2026
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China