

Document No.: rcAAV-BRP-ZL-07

Version No.: A/0

**rcAAV-2/N Quantitation Kit**

Product No.: 1403444

Lot No.: 20251002

Date of Mfg.: 30 Oct. 2025

Expiration Date: 09 Jan. 2027

Test Item	Specification		Result
Linearity of Standard curve	Target	$R^2 \geq 0.990$ .	PASS
	Reference		
Amplification efficiency of Standard curve	Target	E=83.3%-110.0%.	PASS
	Reference		
Repeatability	Target	Low concentration sample was detected, $CV \leq 20.0\%$ .	PASS
	Reference		
No Template Control (NTC)	The Ct value of NTC should be not less than 35.00 or absence of specific amplification.		PASS

Approved By Zhang LipingDAY/MONTH/YEAR 03/11/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China