

Document No.: Recombinant Cascade-BRP-ZL-03

Version No.: A/0

**Recombinant Cascade Reagent (rCR)
(Kinetic-Chromogenic Assay)**Product No.: **1501112**Lot No.: **20251002**Date of Mfg.: **28 Oct. 2025**Expiration Date: **16 Apr. 2026**

Test Item	Specification	Result
Appearance	The package is intact, the label is legible and correct. All components are present as specified. The liquid component is clear and transparent. The freeze-dried preparation appears as a loose, white, block-like solid.	PASS
Linearity of Standard Curve	$ r \geq 0.980$.	PASS
Repeatability	0.5 EU/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Negative Control	The value of the negative control does not exceed the limit of the kit.	PASS

Approved By Cheng JinqinDAY/MONTH/YEAR 30/10/2015
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China