

Document No.: Virus-BRP-ZL-54

Version No.: A/0

REO-3 Detection KitProduct No.: **1506733**Lot No.: **20251101**Date of Mfg.: **17 Nov. 2025**Expiration Date: **22 Apr. 2027**

Test Item	Specification	Result
Low concentration sample	FAM channel: Ct < 38.00 and specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS
No Template Control (NTC)	FAM channel: Ct ≥ 40.00 or absence of specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 17/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China