

Document No.: Virus-BRP-ZL-53

Version No.: A/0

**REO-3 Positive Control**Product No.: **1506737**Lot No.: **20250901**Date of Mfg.: **17 Sep. 2025**Expiration Date: **21 May 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 and specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Ct $\geq$ 40.00 or absence of specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 23/09/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China