

Document No.: Bovine-BRP-ZL-04

Version No.: A/0

Residual Bovine DNA Quantitation KitProduct No.: **1101112**Lot No.: **20251102**Date of Mfg.: **17 Nov. 2025**Expiration Date: **19 May 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or absence of specific amplification.	PASS
Bovine DNA Control	Report the results.	102.9 ng/ μ L

Approved By Zong WeiyiDAY/MONTH/YEAR 24/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China