

Document No.: CHO-BRP-ZL-14

Version No.: A/3

Residual CHO DNA Quantitation Kit (2G)

Product No.: 1101100-1

Lot No.: 20251101

Date of Mfg.: 14 Nov. 2025

Expiration Date: 26 May 2027

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	3 fg/ μ L CHO DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS
CHO DNA Control	Report the results.	30.3 ng/ μ L

Approved By Zong WeiyangDAY/MONTH/YEAR 26/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China