

Document No.: CHO-BRP-ZL-14

Version No.: A/3

Residual CHO DNA Quantitation Kit (2G)Product No.: **1101100-1**Lot No.: **20251201**Date of Mfg.: **03 Dec. 2025**Expiration Date: **24 Aug. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	3 fg/ μ L CHO DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS
CHO DNA Control	Report the results.	30.9 ng/ μ L

Approved By Zong WeijiangDAY/MONTH/YEAR 10/12/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China